

☐ SECRET☒ CONFIDENTIAL☐ UNCLASSIFIED

CONTRACT INSPECTION REPORT		CONTRACT NO. <div style="border: 1px solid black; height: 20px; width: 100%;"></div>	TASK NO.		
TO: CONTRACT ADMINISTRATION & SETTLEMENT BRANCH/PD/OL		DATE 14 December 1965			
		INSPECTION REPORT NO. (If final, so state) 2			
		ESTIMATED COMPLETION DATE 30 December 1965			
NAME OF CONTRACTOR <div style="border: 1px solid black; height: 20px; width: 100%;"></div>					
TYPE OF COMMODITY OR SERVICE Diazo Printer-Processor					
THE CONTRACTOR IS ON SCHEDULE <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO		THE CONTRACTOR WILL PROBABLY REMAIN WITHIN ALLOCATED FUNDS <input checked="" type="checkbox"/> YES <input type="checkbox"/> NO IF ANSWER IS "NO" ADVISE RECOMMENDATION AND/OR ACTION OF SPONSORING OFFICE, ON REVERSE HEREOF. IF KNOWN, INDICATE MAGNITUDE OF ADDITIONAL FUNDS INVOLVED.			
PER CENT OF WORK COMPLETED - 50% PER CENT OF FUNDS EXPENDED - Fixed Price					
HAS AN INTERIM REPORT, FINAL REPORT, PROTOTYPE, OR OTHER END ITEM BEEN RECEIVED FROM THE CONTRACTOR DURING THE PERIOD? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO (If yes, give details on reverse side.)					
HAS GOVERNMENT-OWNED PROPERTY BEEN DELIVERED TO CONTRACTOR DURING THIS PERIOD? <input type="checkbox"/> YES <input checked="" type="checkbox"/> NO (If yes, indicate items, quantity, and cost on reverse side.)					
INCENTIVES					
IS THIS AN INCENTIVE CONTRACT IF YES, CHECK TYPE <input type="checkbox"/> COST <input type="checkbox"/> PERFORMANCE <input type="checkbox"/> DELIVERY		<input type="checkbox"/> YES <input checked="" type="checkbox"/> NO NOTE: USE REVERSE SIDE FOR COMMENTS. FINAL REPORT MUST CONTAIN INCENTIVE EVALUATION.			
OVERALL PERFORMANCE OF CONTRACTOR					
1. <input checked="" type="checkbox"/> OUTSTANDING	3. <input checked="" type="checkbox"/> ABOVE AVERAGE	5. <input type="checkbox"/> BELOW AVERAGE	7. <input type="checkbox"/> UNSATISFACTORY		
2. <input type="checkbox"/> EXCELLENT	4. <input type="checkbox"/> AVERAGE	6. <input type="checkbox"/> BARELY ADEQUATE			
IF OVERALL PERFORMANCE OF CONTRACTOR IS UNSATISFACTORY OR BARELY ADEQUATE, INDICATE REASONS ON REVERSE SIDE.					
RECOMMENDED ACTION					
<input type="checkbox"/> CONTINUE AS PROGRAMMED		<input type="checkbox"/> WITHHOLD PAYMENT PENDING SATISFACTORY PERFORMANCE			
<input type="checkbox"/> TERMINATE		<input type="checkbox"/> OTHER (Specify)			
IF TERMINATION IS RECOMMENDED OR IF THIS IS A FINAL REPORT PUT COMMENTS ON REVERSE IN NARRATIVE FORM ON CONTRACTOR'S PERFORMANCE AND CERTIFY THAT ALL DELIVERABLE ITEMS UNDER THE CONTRACT HAVE BEEN RECEIVED. THESE INCLUDE, WHERE APPLICABLE, THE FOLLOWING:					
ITEM	REC'D	DOES NOT APPLY	ITEM	REC'D	DOES NOT APPLY
PROTOTYPES			MANUALS		
DRAWINGS AND SPECIFICATIONS			FINAL REPORT		
PRODUCTION AND/OR OTHER END ITEMS			SPECIAL TOOLING		
			OTHER GOVERNMENT PROPERTY		
DATE OF LAST CONTACT WITH CONTRACTOR 7 December 1965					
SIGNATURE OF INSPECTOR <div style="border: 1px solid black; height: 40px; width: 100%;"></div>			DIVISION P & D S		
INSPECTOR'S			SIGNATURE OF APPROVER		

25X1

25X1

25X1

NARRATIVE REPORT

☒ INTERIM☐ FINAL

Inspection Visit of 1 December 1965

ATTENDEES:

- 1.
- 2.
- 3.
- 4.
- 5.

25X1

This visit was made for final inspection of the printer/processor, but due to last minute difficulties the instrument did not pass inspection. Specifically, the following items were not acceptable:

1. The instrument would not produce the required resolution or density range.

	Required	Produced
A. Resolution	200 lp/mm	100 lp/mm
B. D _{max}	1.80	1.60
C. D _{min}	>.1	>.1
D. Gray Steps	18	13
E. T _{max}	>20 sec.	>20 sec.

2. ☐ will modify the microscope mount (y-slide) to eliminate interference with the film storage drawer. It is also possible, that the door on the drawer may be modified.
3. The film processor is now mounted on the elevating table frame rather than on the underside of the light table. This will reduce the weight that is elevated when the table is raised or lowered. The processor scratches the diazo prints and this must be corrected.
4. An inspection visit will be made on 16 December to see if the above items have been corrected.

25X1

☐ UNCLASSIFIED☒ CONFIDENTIAL☐ SECRET